



Personal Protective Equipment Regulation (EU) 2016/425

Certificate

Module B EU Type-Examination

Manufacturer

Burlington Medical, LLC

3 Elmhurst Street, Newport News, VA 23603, USA

Product Description:

X-ray protective aprons and associated items

Product Code:

See page 2 for products codes

Technical File Reference:

522609

Harmonised Standard(s):

EN ISO 13688:2013/A1:2021

Technical Specification:

EN 61331-1:2014; EN 61331-3:2014
(Modified Broad Beam Geometry, BBG*)

Certificate Number: SH00018

Issued by: Shirley® (Notified Body No. 2895 for Regulation (EU) 2016/425)

First issue: 22 June 2021

Date of Issue: 14 March 2025

Expiry: 16 March 2030

Authorised by

C A Butcher

Certification Manager

The attached schedule of approval forms part of this certificate.

Note: The validity of this certificate can be confirmed by contacting the Issuing Office:
Shirley Technologies (Europe) Limited, Sky Business Centre, Port Tunnel Business Park, Office 13
Unit 21, Clonsaugh Business & Technology Park, Dublin 17, ROI

Tel: +353 (0) 01894 1448 **email:** info@shirley.ie **website:** www.shirley.ie





Schedule of Approval Certificate Number: SH00018

First Issued: 22 June 2021
Issue date: 14 March 2025
Expiry date: 16 March 2030

Page No: 2 of 6
Issued by: Shirley® (Notified Body No. 2895)
Shirley® ref: SH-038428

Manufacturer: Burlington Medical, LLC
Technical file ref: 522609

Shirley®, specified as a "notified body" under the terms of the Regulation (EU) 2016/425 of the European Parliament and of the Council on personal protective equipment, did undertake the relevant type approval procedures for the equipment identified below which was found to be in compliance with the relevant provisions of Annex V (Module B) of the Regulation and with the applicable essential health and safety requirements, subject to any conditions in the schedule attached hereto.

The certificate relates specifically to the PPE items described and depicted in the manufacturer's Technical File, copies of which are held by the manufacturer and Shirley®, and not to any other items.

The certificate remains valid unless cancelled or revoked, provided the conditions in the attached schedule are complied with and the equipment remains satisfactory in service.

Description of product

X-ray protective aprons and associated items consisting of:

EN 61331-1:2014 & 61331-3:2014

Frontal Aprons	A10; F17; 601; F01; P500
Wrap Around Aprons	F111; F112; W113; W114; P600; Euro
Wrap Around Vest	Vest; PVest; FVest; QVest; Zvest
Wrap Around Kilt/Skirt	K303; K102
Thyroid Collars	TSS; TSO; TSV; TSB; TSC; BAT
Gonad Shield/Half Apron	HALF

EN 61331-1:2014 Only

Sleeves	SL; SL3
Wrap Around Kilt/Skirt	Kilt
Accessories	Cap; Blocker; Kilt Blocker; Leaded Hood; LGW Leg Wraps
Ethereal Shields (0.125 mmPb)	F17; Euro; K303; PVest; QVest; Zvest; Vest; FVest; Kilt; F111; F112
Additional Breast & Axillary protection to be used with thyroid collar:	BAT

CONTINUED ON PAGE 3



Schedule of Approval
Certificate Number: SH00018

First Issued: 22 June 2021
Issue date: 14 March 2025
Expiry date: 16 March 2030

Page No: 3 of 6
Issued by: Shirley® (Notified Body No. 2895)
Shirley® ref: SH-038428

Manufacturer: Burlington Medical, LLC
Technical file ref: 522609

Available in the following materials:

Outer/Inner Fabrics: Dartex / Vinyl / Polyester / Nylon-Polyester Blend / Polyurethane / Patterns

Protective Core Materials:

Cost Cruncher (Lead):
 $W_s \geq 3.49 \text{ kg/m}^2$ with a Lead Equivalence of 0.25 mmPb for 50 – 150 kV
 $W_s \geq 4.83 \text{ kg/m}^2$ with a Lead Equivalence of 0.35 mmPb for 50 – 150 kV
 $W_s \geq 7.01 \text{ kg/m}^2$ with a Lead Equivalence of 0.50 mmPb for 50 – 150 kV

Truelite Mix Lead (Composite):
 $W_s \geq 3.41 \text{ kg/m}^2$ with a Lead Equivalence of 0.25 mmPb for 90 – 110 kV
 $W_s \geq 4.51 \text{ kg/m}^2$ with a Lead Equivalence of 0.35 mmPb for 60 – 110 kV
 $W_s \geq 6.46 \text{ kg/m}^2$ with a Lead Equivalence of 0.50 mmPb for 60 – 110 kV

Scatter Sentry (Lead Free); Lead Free;
 $W_s \geq 2.79 \text{ kg/m}^2$ with a Lead Equivalence of 0.25 mmPb for 70 – 90 kV
 $W_s \geq 3.86 \text{ kg/m}^2$ with a Lead Equivalence of 0.35 mmPb for 70 – 90 kV
 $W_s \geq 5.79 \text{ kg/m}^2$ with a Lead Equivalence of 0.50 mmPb for 60 – 110 kV

Xenolite Strata 300 Bi-layer Lead Free;
 $W_s \geq 2.96 \text{ kg/m}^2$ with a Lead Equivalence of 0.25 mmPb for 60 – 110 kV
 $W_s \geq 4.17 \text{ kg/m}^2$ with a Lead Equivalence of 0.35 mmPb for 60 – 110 kV
 $W_s \geq 5.94 \text{ kg/m}^2$ with a Lead Equivalence of 0.50 mmPb for 60 – 110 kV

Xenolite Strata 500; Lead Free;
 $W_s \geq 1.61 \text{ kg/m}^2$ with a Lead Equivalence of 0.125 mmPb for 50 – 150 kV
 $W_s \geq 3.04 \text{ kg/m}^2$ with a Lead Equivalence of 0.25 mmPb for 50 – 150 kV
 $W_s \geq 4.34 \text{ kg/m}^2$ with a Lead Equivalence of 0.35 mmPb for 50 – 150 kV
 $W_s \geq 6.08 \text{ kg/m}^2$ with a Lead Equivalence of 0.50 mmPb for 50 – 150 kV

Xenolite 200T;
 $W_s \geq 3.16 \text{ kg/m}^2$ with a Lead Equivalence of 0.25 mmPb for 60 – 150 kV
 $W_s \geq 4.27 \text{ kg/m}^2$ with a Lead Equivalence of 0.35 mmPb for 60 – 120 kV
 $W_s \geq 5.96 \text{ kg/m}^2$ with a Lead Equivalence of 0.50 mmPb for 50 – 120 kV

Xenolite 100-L;
 $W_s \geq 3.45 \text{ kg/m}^2$ with a Lead Equivalence of 0.25 mmPb for 50 – 150 kV
 $W_s \geq 5.20 \text{ kg/m}^2$ with a Lead Equivalence of 0.35 mmPb for 50 – 150 kV
 $W_s \geq 6.90 \text{ kg/m}^2$ with a Lead Equivalence of 0.50 mmPb for 50 – 150 kV

CONTINUED ON PAGE 4



Schedule of Approval Certificate Number: SH00018

First Issued: 22 June 2021
Issue date: 14 March 2025
Expiry date: 16 March 2030

Page No: 4 of 6
Issued by: Shirley® (Notified Body No. 2895)
Shirley® ref: SH-038428

Manufacturer: Burlington Medical, LLC
Technical file ref: 522609

Manufacturers Technical Specification

The manufacturer's Technical Specification for the end use of X-Ray Protective Clothing and accessories was based on testing according to EN 61331-1:2014. Product design is based on EN 61331-3:2014 with variations and adaptations based on user requirements.

The suitability of this specification was checked with respect to the Essential Health and Safety Requirements of (EU) Regulation 2016/425 and was found to address the requirements for this end use.

Testing of the core materials was carried out according to the Modified Broad Beam Geometry (BBG*), as detailed in the 'Recommendation for Use' PPE-R/05.34-001 approved by the Co-ordination of Notified Bodies Vertical Group 5.

Limitations of Use

- Usage, maintenance and storage as per manufacturer's instructions.

Observations

- Not applicable.

Authorised Sub Manufacturer

ProtecX Medical Limited; 1-4 Knowl Piece Business Centre, Knowl Piece, Wilbury Way, Hitchin, Hertfordshire, SG4 0TY, UNITED KINGDOM

CONTINUED ON PAGE 5



Schedule of Approval
Certificate Number: SH00018

First Issued: 22 June 2021
Issue date: 14 March 2025
Expiry date: 16 March 2030

Page No: 5 of 6
Issued by: Shirley® (Notified Body No. 2895)
Shirley® ref: SH-038428

Manufacturer: Burlington Medical, LLC
Technical file ref: 522609

Approval Documents

- Lead Equivalence test report Nos. PPb 05/2019; 2023050283-2; 0027-24; 0091-22; 2023060159; 0087-24; 2024040218_1A
- Innocuousness test report No. 44419; AR-24-S6-004096-01

CONTINUED ON PAGE 6



Schedule of Approval Certificate Number: SH00018

First Issued: 22 June 2021
Issue date: 14 March 2025
Expiry date: 16 March 2030

Page No: 6 of 6
Issued by: Shirley® (Notified Body No. 2895)
Shirley® ref: SH-038428

Manufacturer: Burlington Medical, LLC
Technical file ref: 522609

Terms and Conditions associated with the issue of this EU Type-Examination Certificate

1. This certificate is issued subject to Shirley®'s standard terms of business, available from our website.
2. Production is limited to the site(s) listed in the manufacturer's Technical File, copies of which are held by the manufacturer and Shirley®, and not to any other production site(s).
3. The client must implement appropriate changes as notified by Shirley®.
4. The client must ensure the certified product is representative of the ongoing manufactured product.
5. The client must make provision for access to relevant documents and records.
6. The client must investigate complaints associated with the certified products. Records of such complaints, and actions taken, must be kept by the client and made available to Shirley® when requested.
7. The client must only make claims consistent with the scope of certification,
8. The client must not make any misleading or unauthorised comments regarding the certified product or the certification body.
9. The client must upon suspension, withdrawal, or termination of certification discontinue the use of all advertising matter that contains any reference thereto and take action to return this certificate to Shirley®.
10. The client must comply with the requirements for the use of the notified body number as detailed below.
11. Any change to the product or quality manual / quality plan shall be immediately notified to Shirley®.
12. This certificate is issued in the English language only. It is the responsibility of the Manufacturer / Authorised Representative to obtain and supply language versions acceptable to the country where the product is to be sold. This certificate remains the property of Shirley® and will be withdrawn if any of the conditions attached to its issue are not complied with.
13. The EC mark consists of the letters 'CE', in the form given in Annex II of Regulation (EC) No 765/2008 and for category III PPE, followed by the number of the notified body involved in production control monitoring (Module C2 or D).
14. This certificate does not authorise the use of the Mark of Conformity (the 'CE mark'), which may only be affixed to the above type approved equipment and a Manufacturer's Declaration of Conformity issued when Module C2 or D of the Regulation is fully complied with and controlled by a written agreement with a notified body.

Use of Notified Body Number

1. The Notified Body Number must only be used
 - a. In direct association with products or systems covered by this Type-Examination Certificate.
 - b. by holder(s) of the Certificate.
2. Use of Shirley® Notified Body Number does not extend to other companies which are:
 - a. part of the same corporate group as the Certificate holding company: or
 - b. named in a Certificate, for example as a supplier.
3. Particular care must always be taken to avoid the association of the Shirley® Notified Body Number with other products or systems or schemes and with claims or information not contained in the Shirley® document.

If any of the above requirements are not met Shirley® will seek to suspend, withdraw or terminate this certificate.

END OF SCHEDULE